



VMRD

PO Box 502, Pullman, WA 99163 USA

Telephone: + 1 (509) 334-5815

Fax: + 1 (509) 332-5356

E-mail: vmrd@vmrd.com

Web site: <http://www.vmrd.com>

Certificate of Analysis

Equine Infectious Anemia Reference Serum, ELISA Weak Positive

Reference serum


OD Values		Average OD
0.158	0.189	0.174

CATALOG NO.: RS-EIA-EW-0.5ML

VOLUME: 0.5 ml

LOT: P070507003

EXPIRATION DATE: 9 June 2011



SPECIES OF ORIGIN: Equine

DESCRIPTION: Equine serum containing a level of antibody to Equine Infectious Anemia Virus that gives a weak positive reaction in the VMRD ELISA and a negative reaction in the VMRD AGID.

QUALITY CONTROL METHOD: Sample was tested in the VMRD EIA ELISA (catalog no. 290-1) and AGID (catalog no. 400-200) according to each kit insert.

ELISA Results: Average OD 0.174

AGID Results: Negative.

OTHER INFORMATION: The sample's expected OD range in the VMRD EIA ELISA is > 0.1-0.4. It is also usually negative by AGID.

INTENDED USE: This product is intended as a reference sample for quality assurance of EIA ELISA tests.

STORAGE: This reference serum is provided in liquid form and should be stored at 2-8°C. If serum becomes cloudy, it should be discarded. This serum contains 10 ppm ProClin 300 as a preservative.

FOR *IN VITRO* LABORATORY USE ONLY.

WARRANTY: VMRD, Inc. warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. VMRD, Inc. is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.