

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P260527
Expiration:	13 July 2034

Quality Control Method: Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no. NC-IFA-FIP1), and anti-feline IgG affinity purified FITC conjugate (catalog no. CJ-F-FELG-AP-10ML).

Specific Reaction: 3–4+ signal with the positive control and negative, no background with the negative control. There are 0–50 infected cells per high-power field.

Other Comments: NA

QC Manager:



TJ Yockachonis

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

14 July 2026